

CENTRIFUGE EXTRACTORS CE-3000

STANDARD: EN 12697-1 CLAUSE B.1.5, ASTM D2172, AASHTO T164A

The centrifuges are used for the determination of bitumen percentage in bituminous mixtures.

All models comprise a removable precision-machined rotor bowl housed in a cylindrical aluminium box.

The rotating unit is suspended on the base by four calibrated springs, which assure a perfect stability all over the test. The cover is precisely machined and fitted with solvent resistant gasket to avoid leakages. All models are fitted, for emergency use, by a hand brake system. The control panel includes: Start/Stop button, speed control knob.



SINCE 2006

ZHUOZHOU TIANPENG INSTRUMENT MANUFACTURING CO., LTD.

Website: www.testmould.com Tel:+86-312-3852880